

Application note | 002748

ThermoFisher
SCIENTIFIC

Environmental

Screening of PFAS compounds in wastewater using adsorbable organic fluorine with combustion ion chromatography (CIC)

Results from U.S. EPA draft Method 1621 Collaboration Study

Authors

Terri Christison and Neil Rumachik

Thermo Fisher Scientific
Sunnyvale, CA

Keywords

EPA Method 1621, AOF, CIC, PFOA, PFHxS, GAC, Dionex IonPac AS24 column, PFAS (per- and polyfluoroalkyl substances), combustion ion chromatography (CIC), adsorbable organic fluorine (AOF), granular activated carbon (GAC)

Introduction

Per- and polyfluoroalkyl substances (PFAS) is the collective name for synthetic fluorinated compounds. The regulatory definition varies. The Organisation of Economic Co-operation and Development (OECD) defines a PFAS compound as one containing a fully fluorinated methyl group or methylene on an alkane functional group.¹ The more restrictive U.S. Environmental Protection Agency (EPA) definition, based on the degradation properties, states the fluorinated compound must have at least two adjacent carbons with one carbon fully fluorinated and the other at least partially fluorinated (*unit R-CF₂-CF(R')(R'')*, where R, R', and R'' do not equal "H" and the carbon-carbon bond is saturated).² For example, the EPA classification excludes trifluoroacetic acid (CF₃-COOH) as a PFAS compound because of the absence of an adjoining fluorinated carbon. Since the 1940s, many products have been composed of or contain PFAS. It is found in many polymers, surfactants, fire-suppression chemicals, industrial products, and consumer products.³⁻⁶ In addition to their extensive presence, PFAS compounds are chemically stable due to their C-F bonds, and therefore, they are persistent and bioaccumulate. Consequently, PFAS compounds

are an environmental contamination concern. Extensive toxicity studies of perfluorooctanoic acid (PFOA) and perfluorooctane sulfonate (PFOS) indicate varying toxicity effects from acute to chronic, potentially impacting reproductive health and causing epigenetic effects.⁷⁻¹¹ Less is known about the thousands of other compounds grouped under the EPA PFAS classification.²

The EPA has defined PFAS analysis methods for select PFAS analytes using LC-MS/MS in drinking water (EPA Methods 533.0 and 537.1) and in aqueous, solid, biosolids, and tissue samples (EPA Method 1633, at the publishing date of this document in the 90-day comment phase prior to promulgation).^{12,13} Thermo Fisher Scientific has also previously demonstrated these methods.¹⁴⁻¹⁸ However, a screening method is needed to assess the extent of PFAS contamination in various matrices, including wastewater. Pyrohydrolytic combustion-ion chromatography (CIC) has been previously demonstrated to eliminate the sample matrices.^{19,20} More recently, CIC has been proposed as a screening method for PFAS by converting all organofluorine compounds (including PFAS) to HF, and the subsequent fluoride analyzed by ion chromatography with suppressed conductivity detection (IC).²¹⁻²⁷ The draft EPA Method 1621 has completed a single laboratory review and a round-robin collaborative study and was issued as an official method in the 90-day review and promulgation stage at the time of publication of this document.²⁸⁻³⁰ In this method, 100 mL of wastewater are adsorbed onto granular activated carbon (GAC) columns. Inorganic fluoride is rinsed from the GAC columns with 10 mM sodium nitrate followed by a deionized (DI) water rinse. The GAC is removed from the glass tube, combusted under pyrohydrolytic conditions, absorbed in DI water, and analyzed for fluoride by IC. In this application note, our results from the collaboration study are presented alongside practical advice that supports successful analyses. EPA draft Method 1621 includes method validation tests, calibration, method blanks, method detection limits, multiple process check standards, analysis of wastewater samples, and recovery results of four types of PFAS (PFHxS, PFOS, PFOA, and a blend described as a mixture of PFAS compounds).

Experimental

Equipment

Any Thermo Scientific™ Dionex™ RFIC systems with a conductivity detector (CD) combined with any combustion-absorption system and offline adsorption system can be used.

Recommended: Thermo Scientific™ Dionex™ Inuvion™ IC system with RFIC ([P/N 22185-60108](#)), Thermo Scientific™ Dionex™ Integrion™ RFIC system ([P/N 22153-60305](#)), and Thermo Scientific™ Dionex™ ICS-6000 HPIC™ system.

The Dionex ICS-6000 HPIC system was used in this study, including

- SP Single Pump, isocratic ([P/N 22181-60003](#))
- EG Eluent Generator ([P/N 22181-60019](#))
- DC Detector Column module ([P/N 22181-60043](#))
- CD Conductivity Detector ([P/N 079829](#))

In this study, Nittoseiko Analytech (Nittoseiko Analytech Co., Ltd) AQF-2100H combustion-absorption system with an autosampler for solid samples and an offline adsorption unit was used.

Software

Thermo Scientific™ Chromeleon™ Chromatography Data System (CDS) software, version 7.3.1 or higher, was used to control the IC. Contact your local Thermo Fisher Scientific specialist for the latest information on combustion unit control.

Consumables list for EPA Method 1621

- Thermo Scientific™ Dionex™ IC PEEK Viper™ fitting kits for:
 - Dionex Inuvion RFIC system ([P/N B51000232](#))
 - Dionex Integrion RFIC system with CD detection ([P/N 088798](#))
 - Dionex ICS-6000 HPIC system with CD detection ([P/N 302965](#))
- Thermo Scientific™ Dionex™ EGC 500 KOH eluent generator cartridge ([P/N 075778](#))
- Thermo Scientific™ Dionex™ CR-ATC 600 electrolytic trap column ([P/N 088662](#))
- Thermo Scientific™ Dionex™ HP EG degasser kit ([P/N 075522](#))
- Thermo Scientific™ Dionex™ ADRS 600 suppressor (2 mm) ([P/N 088667](#))
- Thermo Scientific™ Dionex™ IonPac™ AS24 analytical column ([P/N 064153](#))

- Thermo Scientific™ Dionex™ IonPac™ AG24 guard column (P/N 064151)
- Hamilton™ 100 µL microsyringe, metal plunger, Fisher Scientific (P/N 14-813-091)
- Fisherbrand™ Elite™ adjustable, 200 µL pipette and pipette tips, Fisher Scientific (P/N FBE00200 and P/N 02-707-409)
- HDPE 1.0 mL vials and caps, Fisher Scientific (P/N NC2038456, P/N 6ERC11PEC)
- Ceramic combustion boats for Nittoseiko Analytec ASC-270LS sample changer, Nittoseiko Analytec TX3SCY
- Granular activated carbon (GAC) columns used to adsorb samples in EPA Method 1621 multi-lab collaboration study with corresponding GAC column carriers, Analytik Jena 402-880.615, 402-880.013
- Quartz wool used in combustion tube and used to accumulate solids in high TSS samples, Nittoseiko Analytec TNQWL5K

Method

The method includes adsorption, combustion-absorption, and analysis by ion chromatography.

Adsorption conditions

All standards and samples are processed through the adsorption process discussed in section 11.3, EPA Method 1621. The empty boats, calibration standards, and calibration verification standards (CV) are the exceptions. They skip the adsorption process and enter the workflow at the combustion stage. For CV and calibration standards, pipette 200 µL of the individual µg/mL (mg/L) standard into a clean ceramic boat (resulting in “ng” amounts of fluoride). For example, 200 µL of 2.5 µg/mL fluoride = 500 ng.

Parameter	Setting
Adsorption	Pump 100 mL of each sample or standard onto the GAC columns.
Rinse off inorganic fluoride	Pump 25 mL 10 mM sodium nitrate for each sample, MB, Initial Precision Recover (IPR) standard, Ongoing Precision Recovery standard (OPR), or blank ceramic wool onto the GAC columns.
Rinse off nitrate	Pump 20 mL of DI water onto the GAC columns or blank ceramic wool.
Dry	Pump 5 mL of air onto the GAC columns or blank ceramic wool.

Combustion and absorption conditions

Parameter	Setting		
Transfer samples	Transfer GAC material from each GAC column or quartz wool from the GAC carrier into a separate ceramic boat. Combust according to the boat program.		
Boat program	<i>Position</i>	<i>Wait time (s)</i>	<i>Speed (mm/s)</i>
	90 mm	20	10
	End	600	10
	Cool	60	40
	Home	120	20
Furnace temp.	950 °C inlet, 1,000 °C outlet		
Combustion time	Ar: 10 s; O ₂ : 600 s; Ar replacement: 30 s		
Absorption sol.	3 mL DI water		
Final absorption sol.	10 mL, mixed		
Gas	Ar: 200 mL/min; O ₂ : 400 mL/min		
Hydration	Water/Ar: 100 mL/min		
Injection volume	100 µL		

IC conditions

Parameter	Setting
Columns	Dionex IonPac AG24 guard, Dionex IonPac AS24 analytical anion-exchange columns, 2 mm i.d. Note: Other hydroxide selective columns can be used instead (e.g., Dionex IonPac AS19, AS20, AS30 columns).
KOH gradient***	8 mM KOH (0 to 3.5 min, curve 5), 8–75 mM (3.5 to 9 min, curve 9), 75 mM (9 to 10.25 min, curve 5), 75–8 mM (10.25 to 15 min, curve 5), 8 mM (15 to 20 min, curve 5)
Eluent source	Dionex EGC 500 KOH eluent generator cartridge, Dionex CR-ATC 600 trap column, Dionex HP EG degas module
Injection vol.	100 µL
Flow rate	0.30 mL/min
Column temp.	30 °C
Detection temp.	35 °C
Suppressor temp.	20 °C
Detection	Suppressed conductivity, Dionex ADRS 600 suppressor, recycle mode, 56 mA
Conductance background	<0.5 µS/cm
System backpressure	≈2,600 psi (≈18,000 kPa)
Run time	25 min for the first sample, 20 min in overlap for the following samples

***Curve 5 is linear; Curve 9 is convex

Reagents

- DI water ASTM Type 1 (18 MΩ·cm resistivity, TOC < 50 ng/mL)³¹
- Ammonium hydroxide, 28–30 w/w%, Certified ACS, Fisher Scientific™ [P/N A669S-500](#) (CAS 1336-21-6)
- Methanol, UHPLC-MS grade, Fisher Scientific [P/N A458-1](#) (CAS 67-56-1)
- Sodium nitrate, ACS Certified, Fisher Scientific [P/N S343-3](#) (CAS 7631-99-4)
- 1,000 mg/L chloride standard for retention time determinations, TraceCERT™, MilliporeSigma™ Supelco™ [P/N 1.19897.0500](#), Fisher Scientific [P/N 11-101-6049](#)
- 1,000 mg/L certified fluoride standard for IC, SPEX™ CertiPrep™, Fisher Scientific [P/N AS-F9-2Y](#) or Dionex Sodium fluoride 1,000 mg/L, NIST traceable, [P/N 037158](#)
- 1,000 mg/L Fluoride solution, Ricca Chemical™, used as a second fluoride stock standard, Fisher Scientific [P/N 3173-6](#) or Dionex Sodium fluoride 1,000 mg/L, NIST traceable, [P/N 037158](#)
- 1,000 ng/mL 30-PFAS mixed spiking standard in methanol (Wellington Laboratories [P/N PFAC30 PAR](#), Lot PFAC300522), (Fluoride equivalence = 18.2 ng/mL)
- 50 µg/mL (in methanol) L-PFOS sodium perfluoro-1-octanesulfonate standard (Wellington Laboratories [P/N L-PFOS](#), Lot LPOS0922), used for spike-recovery experiments (Fluoride equivalence = 61.7%; 30.3 µg/mL fluoride)
- 50 µg/mL (in methanol) L-PFHxS sodium perfluoro-1-hexanesulfonate standard (Wellington Laboratories [P/N L-PFHxS](#), Lot LPFHxS1022), used for spike-recovery experiments (Fluoride equivalence = 64.6%; 29.26 µg/mL fluoride)
- 50 µg/mL (in methanol) L-PFBA perfluoro-n-butanoic acid standard (Wellington Laboratories [P/N PFBA](#), Lot PFBA1022), used for spike-recovery experiments (Fluoride equivalence = 62.1%; 30.45 µg/mL fluoride)

Preparation of solutions, standards, and samples

Use only ASTM Type I deionized water (DI water) for preparations.

Rinse all glassware, HDPE containers/bottles, and tools with methanol, followed by DI water, to eliminate potential fluoride contamination.

Adsorption solutions: 2 M sodium nitrate stock solution, 10 mM sodium nitrate wash solution, and 1% ammonium hydroxide in Fisher Chemical™ Optima™ LC-MS grade methanol cleaning solution, as described in section 7.1 of EPA Method 1621.³⁰

Standards, blanks, and samples: Follow EPA Method 1621 instructions to prepare inorganic fluoride and chloride standards. Prepare inorganic fluoride and chloride retention time standards, fluoride CV standards, and fluoride calibration standards (0.5, 1, 2.5, 5, 10, 12.5, 25, 37.5, and 50 µg/mL) from 1,000 mg/L stock standards. Refer to TN003056 for detailed processes to achieve low background levels and minimize cross-contamination needed for this application.³² Follow EPA Method 1621, section 11.3 to process the samples through the adsorption step.

PFAS standards in methanol: Use a microsyringe without Teflon™ or fluoropolymers-containing components to dose PFHxS aliquots or PFAS standards for MDL, IPR, OPR, and recovery samples. Clean the microsyringe multiple times with suitable high-purity methanol. Clean between sample- or standard-spiking. Store standards in fluorocarbon-free vials with lids, such as HDPE, at 20 °C. See EPA Method 1621, section 7.3.3 to calculate total F in the PFHxS standard.

Instrument setup and installation

Potential sources of contamination

To exploit the full potential of the trace analysis method, potential sources of contamination must be excluded in advance. A detailed discussion on achieving low contamination levels and tips for successful implementation of EPA Method 1621 are discussed in TN003056.³²

Adsorption standalone instrument

Set the dials of the instrument to the desired volume. Install empty GAC carriers. Clean the fluid flow with the GAC carriers in position using 1% ammonium hydroxide in methanol rinse followed by DI water rinse.

CIC system

Install and set up the Dionex HPIC system as described in Figure 1 and the instrument manuals.^{33–36} TTL triggers are effective ways to trigger the Dionex IC to start after the combustion-absorption is complete. Refer to the instrument manuals. Thermo Fisher Scientific Technical Note 000767 describes the configuration of CIC with a Dionex Integrion HPIC system using Chromeleon 7 CDS software, which is similar for other Dionex HPIC RFIC systems, such as the Dionex Inuvion RFIC system and the Dionex ICS-6000 HPIC system).³⁷ Hydrate and condition the IC consumables as described in TN000767.




Figure 1. Flow diagram for CIC system

Results and discussion

This application follows EPA draft Method 1621 (similar to EPA Method 1621³⁰) statement of work and technical directives. The collaboration study was organized by the managing laboratory, General Dynamics Information Technology company, which provided GAC columns and GAC column carriers, and PFAS, chloride and fluoride stock standards.³⁸ Definitions are tabulated in the appendix.

Determining the fluoride retention time and window

Fluoride must be well-resolved from the water dip (void volume) and other peaks eluting near fluoride for accurate determinations. In hydroxide eluent separations with suppressed conductivity, as shown in Figure 2, fluoride is well resolved from the water dip, whereas carbonate eluents can be challenging. The results show that fluoride elutes well after (3.3 min) the water dip. The water dip height is also small ($<0.04 \mu\text{S}/\text{cm}$), characteristic of hydroxide eluents. Additionally, after being suppressed, electrolytically generated and purified hydroxide eluents provide higher signal-to-noise ratios, a measure for increased sensitivity and direct compatibility with gradient elution. Fully automated RFIC Eluent Generation (EG) is known for improved analytical and chromatographic performance. The precise and user-independent gradient formation produces superior separations when paired with high-performance Dionex IonPac columns. All the user has to provide is DI water.

Table 1. Determining fluoride retention time window

Peak	Retention time (min)	Peak window (min)	Separation (min)	Limit
Water dip	2.75 ± 0.008	--	--	--
1,250 ng Fluoride*	6.07 ± 0.014	5.98–6.14	3.32	>1 min
5,000 ng Chloride**	12.48 ± 0.005	12.45–12.51	6.42	>1 min
Performance	PASS			

*0.1 mL of 12.5 µg/mL fluoride, n = 3; **0.1 mL of 50 µg/mL chloride, n = 3




Figure 2. Determining fluoride retention time window

Calibration

Calibration curves were determined according to EPA Method 1621 Step 10.2.2, by the peak area response of fluoride versus concentration, in nanograms, after combusting 200 µL of each mg/L calibration standard in separate ceramic boats. The method requires that the best calibration fit is determined by using a relative standard error (RSE) defined in section 10.3.3 Instrument linearity, EPA Method 1621.³⁰

$$RSE = 100 \times \sqrt{\sum_{i=1}^n \frac{[x'_i - x_i]^2}{x_i^2}} \div n - p$$

with

x_i = Nominal concentration (true value) of each calibration standard

x'_i = Measured concentration of each calibration standard

n = Number of standard levels in the curve

p = Type of curve (2 = linear, 3 = quadratic)

Equation 1. Definition of the relative standard error to determine instrument linearity. The RSE must be $\leq 20\%$.

To determine the RSE, evaluate linear and quadratic fits by allowing offset from zero and calculating the curve fit without weighting and with weighting (1/A, 1/A²) by area. Select the calibration curve with the lowest RSE. In our experiments, we used a linear fit with 1/A² area weighting, providing the lowest RSE (0.774) and compliance with the RSE limit of 20% set by the EPA (Table 2). Figure 3 shows the resulting calibration plot.

Initial demonstration of capability

The collaboration study and Section 9.2, EPA Method 1621 require an initial capability demonstration (IDC) before analyzing wastewater samples. The IDC demonstrates that the instruments' and laboratory's performances suit this method. The accuracy (IDA) and precision (IDP) were determined by preparing and analyzing four replicates of DI water fortified with 15 µg/L fluorine (PFHxS).

Table 2. Comparison of relative standard error (RSE) and coefficient of determination (r^2)

	Linear			Quadratic			RSE limit
	No weighting	1/A	1/A ²	No weighting	1/A	1/A ²	
RSE ⁺	24.8	6.01	0.774	4.09	2.05	1.49	20
r^{2++}	0.9991	0.9993	0.9992	0.9996	0.9998	0.9997	

$n = 7$

⁺RSE = Relative Standard Error, weighting = 1/A or 1/A²

⁺⁺ r^2 = Coefficient of Determination

Type: linear with offset, 1/A² weighting




Figure 3. Calibration curve

In contrast to sections 9.2 and 10, the collaboration study required seven replicate method blanks (MB) and 5 µg/L method detection limit (MDL) standards to be prepared and analyzed over three days. The IDA and IDP results demonstrate that the method and this laboratory easily meet the IDC requirements (Table 3). Figure 4 shows the chromatography of the four replicate standards.

Table 3. Initial demonstration of capability

IDC Requirement (Section from EPA Method 1621)	Specification	Experimental value
IDA (9.2.1)	Recovery 80% to 120%	Recovery = 98.5%
IDP (9.2.1)	RSD < 20	RSD = 4.6




Figure 4. Determining initial precision and recovery

After three days of determining MBs and MDLs, the calculated MB was 1.23 ng/mL, and the calculated MDL was 2.5 ng/mL (Table 4). The chromatograms of MB and MDL are compared in Figure 5.

Table 4. Initial demonstration of determining method blanks and method detection limit standards

Day	Method blank Fluoride (ng/mL)	5 ng/mL MDL standard Fluoride (ng/mL)
1	0.51	3.99
1	0.89	4.35
2	0.64	5.24
2	0.67	4.39
3	0.43	5.77
3	0.85	3.88
3	0.81	5.70
Average, X	0.69	4.76
Standard Deviation (σ)	0.17	0.80
RSD	25.5	16.7
MDL	MDL _(b) [^] = 1.23	MDL _(s) ^{^^} = 2.50

$^{\wedge}MDL_{(b)} = X + (3.14 \cdot \sigma)$
Where $MDL_{(b)}$ is the MDL for MBs, X is average of MB, 3.14 is Student's *t*-test value for *n* = 7, 99 percentile, and σ is the standard deviation,³⁷

$^{\wedge\wedge}Section\ ii: MDL_{(s)} = 3.14 \cdot \sigma$
Where $MDL_{(s)}$ is the MDL for spiked MDL standards, 3.14 is Student's *t*-test value for *n* = 7, 99 percentile, and σ is the standard deviation,³⁹




Figure 5. Determining method blank and MDL

Samples

The method was applied to wastewater samples. The results are shown in Table 5. Less than 15 ng/mL of fluoride are present in the samples, ranging from <1 (Sample 7) to 12 ng/mL (Sample 3). The method blanks and process checks show that the method is working well within the expected limits.

Table 5. Sample results and recovery results of added PFHxS standard

Sample	Total fluoride		Recovered (%)
	Measured (ng/mL)	Added (ng/mL)	
Sample 1	9.69	30.0	87.7
		30.0	87.1
		60.3	99.0
		60.3	95.9
Sample 2+	1.11	10.03	116
		10.03	117
		25.2	98.8
		25.2	90.4
Sample 3+	12.22	30.0	92.6
		30.0	91.2
		60.3	87.5
		60.3	88.7
Sample 4	9.69	30.1	108
		30.1	105
		60.3	90.2
		60.3	102
Sample 5	1.56	30.0	103
		30.0	102
		60.3	106
		60.3	104
Sample 6+	4.93	10.09	80.2
		10.09	81.9
		30.0	107
		30.0	105
Sample 7	0.53	10.03	104
		10.03	105
		30.1	102
		30.1	101
Sample 8	1.71	10.03	99.1
		10.03	103
		30.1	92.4
		30.1	96.8
Sample 9	3.23	30.1	98.3
		30.1	96.6
		60.2	85.5
		60.2	95.3

After MBs were subtracted

+ Required ceramic wool

Determining recoveries of added PFHxS standard

To understand method accuracy, recoveries of different volumes of added PFHxS standard were determined at two concentrations and duplicated for all nine samples, including those requiring ceramic wool (Table 5).

The method is accurate, as indicated by the recoveries of added PFHxS standards, 80 to 117%. The chromatogram of Sample 3 is compared with the replicate recovery samples (Figure 6).

Sample prep.: 100 mL adsorbed onto 2 GAC columns
Standards: Chromatograms of only top GAC column
A: Sample 3 wastewater (high TSS)
B: Sample A + 30 ng/mL fluoride equivalence from PFHxS
C: Sample B replicate




Figure 6. Determining recovery of PFHxS in Sample 3 wastewater

Determining recoveries of PFAS compounds in Sample 7

To evaluate the accuracy of determining other PFAS compounds, similar recovery experiments were conducted with added standards of PFBA, PFOS, and a PFAS mixture (Wellington Laboratories). Approximately 10 ng/mL and 30 ng/mL fluoride from the above standards were added to Sample 7 (defined by the managing laboratory). The results summarized in Table 6 show that the fluoride recoveries of PFBA, PFOS, and PFAS mixture ranged from 76 to 114% of added fluoride as perfluorocarbon. Figure 7 shows representative chromatograms of samples after the addition of PFOS, PFBA, and PFAS mixture to determine compound specific recovery of fluoride after combustion. No compound specific dependency was observed.

Table 6. Summary of recovery results from added PFBA, PFOS, and PFAS

Sample	Total fluorine		Recovered (%)
	Added (ng/mL)	Measured (ng/mL)	
Sample 7	–	0.53	–
Sample 7 with PFBA	10.05	10.86	103
	10.05	10.63	100
	30.4	33.31	108
	30.4	23.53	76.1
Sample 7 with PFOS	10.0	10.19	96.8
	10.0	9.08	86.2
	30.3	29.16	94.6
	30.3	35.2	114
Sample 7 with PFAS mix	9.99	11.05	105
	9.99	8.98	85.4
	30.0	25.60	83.9
	30.0	24.91	81.6

Sample prep.: 100 mL adsorbed onto 2 GAC columns
 Samples: Chromatograms of only top GAC column
 A: Sample 7 wastewater.
 Replicates: B) + 10.05 ng/mL fluoride equivalence (PFOS),
 C) + 30 ng/mL fluoride equivalence (PFAS mix),
 D) + 30 ng/mL fluoride equivalence (PFBA)

Peak: 1. Fluoride 0.53 9.08 25.6 33.3 ng/mL
 (86.2%) (83.9%) (108%)




Figure 7. Determining recovery of different PFAS compounds. No compound specific dependency is observed.

Conclusion

This application note demonstrated our results from the multi-laboratory collaboration study of EPA draft Method 1621. This method shows that CIC is an excellent tool to measure AOF and screen for PFAS in wastewater. Along with its ease of use, RFIC provides more accurate results as fluoride is well separated from the water dip, ensuring optimized automated peak integration. The method is sensitive, with MBs of 1 ng/mL and MDLs of 2.5 ng/mL, and accurate, with recoveries of 80–120%. Together, these data highlight the power of CIC in eliminating the sample matrix and measuring only the adsorbable organically bound fluorine content in samples, successfully achieving the goals outlined in EPA draft Method 1621.

References

1. OECD, *Reconciling Terminology of the Universe of Per- and Polyfluoroalkyl Substances: Recommendations and Practical Guidance*, OECD Series on Risk Management, No. 61, OECD Publishing, Paris, France, 2021.
2. European Fluorocarbons Technical Committee (EFCTC), *U.S. EPA working definition for PFAS excludes TFA*, EFCTC, Brussels, Belgium, November 2, 2021.
3. National Toxic Programs (NTP). *Monograph on Immunotoxicity Associated with Exposure to Perfluorooctanoic acid (PFOA) and Perfluorooctane Sulfonate (PFOS)*. Office of Health Assessment and Translation Division of the National Toxicology Program, National Institute of Environmental Health Sciences, National Institutes of Health, U.S. Department of Health and Human Services, **2016**. [viewed Jan 18, 2024]
4. Organisation for Economic Co-operation and Development (OECD), Environmental Directorate Chemicals and Biotechnology Committee. *Fact Cards of Major Groups of Per- and Polyfluoroalkyl Substances (PFASs)*, Series on Risk Management No. 68. ENV/CBC/MONO(2022), Paris, France.
5. OECD. *PFAS AND ALTERNATIVES IN FOOD PACKAGING (PAPER AND PAPERBOARD): HAZARD PROFILE*. Series on Risk Management No. 69. ENV/CBC/MONO(2022)1. Paris, France.
6. McCoy, J. *Is There a 'Right' Definition of PFAS?* I SPH (bu.edu), Boston University, Public Health. March 8, 2022.
7. International Agency for Research on Cancer. *IARC Monographs on the Identification of Carcinogenic Hazards to Humans, Volume 110: Perfluorooctanoic Acid, Tetrafluoroethylene, Dichloromethane, 1,2-Dichloropropane, and 1,3-Propane Sultone*, **2014**.
8. International Agency for Research on Cancer. *IARC Monographs on the Identification of Carcinogenic Hazards to Humans, Volume 135: Perfluorooctanoic acid and perfluorooctanesulfonic acid*. December 2023.
9. Petrof, R.L., Cavalcante, R.G., Langen, E.S., Dolinoy, D.C., Padmanabhan, V., and Goodrich, J.M. Mediation effects of DNA methylation and hydroxymethylation on birth outcomes after prenatal per- and polyfluoroalkyl substances (PFAS) exposure in the Michigan mother–infant pairs cohort. *Clinical Epigenetics*, **2023**, 15:49. doi: <https://doi.org/10.1186/s13148-023-01461-5>
10. Kim, S., Thapar, I., Brooks, B.W. Epigenetic Changes by Per- and Polyfluoroalkyl Substances (PFAS), *Environmental Pollution*, 2021, 279, 15 June, 116929. doi: <https://doi.org/10.1016/j.envpol.2021.116929>

11. Braun, J. Brown epidemiologist breaks down 'forever chemicals' and the research on their health effects | Brown University. Brown University, Health & Medicine topics, July 28, 2022.
12. U.S. Environmental Protection Agency (EPA). Method 537.1 Determination of Selected Per- and Polyfluorinated Alkyl Substances in Drinking Water by Solid Phase Extraction and Liquid Chromatography/Tandem Mass Spectrometry (LC/MS/MS) | Science Inventory | US EPA, EPA/600/R-20/006, U.S. Environmental Protection Agency Office of Water (4303T), Office of Science and Technology Engineering and Analysis Division, Washington, DC. **2020**.
13. U.S. Environmental Protection Agency (EPA). 4th Draft Method 1633 Analysis of Per- and Polyfluorinated Alkyl Substances (PFAS) in Aqueous, Solid, Biosolids, and Tissue Samples by LC-MS/MS (epa.gov), EPA #821-D-23-00, U.S. Environmental Protection Agency Office of Water (4303T), Office of Science and Technology Engineering and Analysis Division, Washington, DC. **2023**.
14. Tracy, M.L. Technical Note 73 Determination of Perfluorooctanoic Acid (PFOA) and Perfluorooctanesulfonic Acid (PFOS) in Water Samples Using On-Line Sample Concentration, Reversed-Phase Liquid Chromatography, and Suppressed Conductivity Detection, **2016**. Thermo Fisher Scientific, Sunnyvale, CA, USA.
15. Zhang, X., Qiu, C., Ullah, R., and Liu, Y. Application Note AN73883, Determination of per- and polyfluorinated alkyl substances (PFAS) in drinking water using automated solid phase extraction and LC-MS/MS for U.S. EPA Method 533, **2020**. Thermo Fisher Scientific, Sunnyvale, CA, USA.
16. Qiu, C., Zhang, X., Ullah, R., Chen, W., and Liu, Y. Application Note: AN73346 Determination of Per- and Polyfluorinated Alkyl Substances (PFAS) in Drinking Water using Automated Solid-Phase Extraction and LC-MS/MS, **2020**. Thermo Fisher Scientific, Sunnyvale, CA, USA.
17. McHale, K. Application Note AN002348 Quantitation of per- and polyfluoroalkyl substances (PFAS) in aqueous samples by LC-MS/MS following EPA Draft Method 1633, **2023**. Thermo Fisher Scientific, Somerset, NJ, USA.
18. Grim, C.¹, McHale, K.², and Astill, T.¹ Application Note AN002771 LC-MS/MS analysis of per- and polyfluoroalkyl substances (PFAS) in soil samples in accordance with EPA Method 1633. Featuring semi-automated solid phase extraction cleanup. **2024**. Thermo Fisher Scientific, ¹San Jose, CA, USA, ²Somerset, NJ, USA.
19. Christison, T.¹, Agustin, A.², and Rohrer, J.¹ Application Note AN73105 Determination of total fluorine, chlorine, bromine, and sulfur in liquefied petroleum gas by pyrohydrolytic combustion ion chromatography, **2019**. ¹Thermo Fisher Scientific, Sunnyvale, CA, USA, ²COSA Xentaur, Houston, TX, USA.
20. Christison, T. and Rohrer, J. Application Note 73865 Pyrohydrolytic combustion ion chromatography: Determination of total chlorine and sulfur in cleanroom gloves, **2020**. Thermo Fisher Scientific, Sunnyvale, CA, USA.
21. Lord, H, Khachatryan, L., Wilson, A. Total Organofluorine (TOF) Analysis by Combustion Ion Chromatography: A New Tool for Monitoring PFAS Impacts. Bureau of Veritas poster. Presented at National Environmental Monitoring Conference, (NEMC), August 1-5, 2022, Crystal City, VA, USA.
22. Miyake, Y., Yamashita, N., Rostkowski, P., So, M.K., Taniyasu, S., Lam, P.K.S., Kannan, K. Determination of trace levels of total fluorine in water using combustion ion chromatography for fluorine: A mass balance approach to determine individual perfluorinated chemicals in water. *J. Chrom A*, **2007**, 1143, 98–104. <https://doi.org/10.1016/j.chroma.2006.12.071>
23. Hu, J.; Rohrer, J. Application Note 72333 Determination of Adsorbable Organic Halogen in Wastewater Using a Combustion Ion Chromatography System, **2017**. Thermo Fisher Scientific, Sunnyvale, CA, USA.
24. Neist, U., Klocke, I., Georgii, S. PhD., Brunn, H., Prof. PhD., and Jensen, D. Customer Application Note CAN 73481 AOF by combustion IC – non-targeted complementary determination of PFAS in aqueous samples. **2020**. Hessian State Laboratory, Wiesbaden, Germany and Thermo Fisher Scientific, GmbH, Dreieich, Germany.
25. Impellitteri, C. Session 2: Analytical methods for PFAS in environmental media, PFAS Science Webinars for EPA Region 1 and State & Tribal Partners. September 16, **2020**.
26. Han, Y., Pulikkal, V.F., and Sun, M. Comprehensive Validation of the Adsorbable Organic Fluorine Analysis and Performance Comparison of Current Methods for Total Per- and Polyfluoroalkyl Substances in Water Samples. *ACS EST Water*, **2021**, 1, 1474–1482. DOI: 10.1021/acsestwater.1c00047
27. Jones, J.L., Burket, S.R., Hanley, A., Shoemaker, J.A. Development of a standardized adsorbable organofluorine screening method for wastewaters with detection by combustion ion chromatography. *Analyt. Methods*, **2022**, 36.
28. U.S. EPA Environmental Protection Agency, Office of Water. Report on the Single-laboratory Validation of Clean Water Act Method 1621 for Adsorbable Organic Fluoride (AOF), EPA 820-R-22-003, April 2022.
29. U.S. Environmental Protection Agency (EPA). Multi-laboratory Validation Study Report for Method 1621: Determination of Adsorbable Organic Fluorine (AOF) in Aqueous Matrices by Combustion Ion Chromatography (CIC), EPA 821-R-24-003. U.S. Environmental Protection Agency, Office of Water (4303T), Office of Science and Technology Engineering and Analysis Division, Washington, DC, USA, January 2024.
30. U.S. Environmental Protection Agency (EPA). Method 1621 Determination of Adsorbable Organic Fluorine (AOF) in Aqueous Matrices by Combustion Ion Chromatography (CIC), EPA 821-R-24-002, U.S. Environmental Protection Agency, Office of Water (4303T), Office of Science and Technology Engineering and Analysis Division, Washington, DC, USA, January 2024.
31. ASTM International, ASTM D1193 - 99e1 Standard Specification for Reagent Water. <https://www.astm.org/d1193-06r18.html>
32. Thermo Fisher Scientific Technical Note 003056 Techniques for successful implementation of EPA Method 1621: Screening of PFAS compounds using adsorbable organic fluorine with combustion IC. Sunnyvale, CA, USA, 2024.
33. Thermo Fisher Scientific: Dionex ICS-6000 operator's manual
34. Thermo Fisher Scientific: Dionex Integrion operator's manual
35. Thermo Fisher Scientific: Dionex Inuvion operator's manual.
36. Mitsubishi Analytec (Nittoseiko Analytec) AQF-2100 Automated Combustion System, Operator's manual.
37. Thermo Fisher Scientific: Technical Note 00767 Combustion ion chromatography with a Dionex Integrion HPIC system using Chromeleon 7 CDS software. Sunnyvale, CA, USA, **2022**.
38. General Dynamics Information Technology, Company. Statement of Work (SOW) for Multi-laboratory Validation of Draft EPA Method 1621 for Adsorbable organic Fluorine by Combustion Ion Chromatography; Technical Directives. September 2022.
39. U.S. Federal Register: 40CFR136, Appendix B, Sections ii, and iii.

Appendix

Table A1. Definitions

Term	Definition	Description
AOF	Adsorbable Organic Fluorine	Fluorocarbon and perfluorocarbon compounds in water or solvent that adhere to a sorbent like activated carbon or ion exchangers.
CV	Calibration Verification standard	Check standards prepared from an inorganic fluoride standard that are from a different (second) source than the calibration standards. Added directly to combustion cups. Run at the beginning and the end of each sequence.
GAC	Granular Activated Carbon	Carbon material used for adsorbing organic compounds.
IPR	Initial Precision Recovery standard	Used to determine the expected process check value (OPR). Prepared by adding PFHxS standard to DI water. Processed through adsorption, combustion, absorption, and IC analysis.
MB	Method blanks	DI water processed through adsorption, combustion, absorption, and IC analysis. Indicates the lowest actual measurements possible and the baseline contamination of the CIC and the Filtration module. Always run at the beginning and the end of each sequence.
MDL	Method detection limit	The lowest reliable measurement, typically 3× S/N. In this test, the standard is processed through the adsorption on carbon, combusted, absorbed, and analyzed by IC.
MDL_(b)	Method detection limit of blanks, n = 7	Defined as: (Mean of method blanks) plus (standard deviation × Student <i>t</i> -test factor).
MDL_(s)	Method detection limit of 5 ng/mL PFHxS standard, n = 7	Defined as: (standard deviation) × (Student <i>t</i> -test factor).
OPR	Ongoing Precision Recovery standard	Check standard run at the beginning of each sequence. Prepared by adding PFHxS standard to 100 mL of DI water. Processed through the adsorption on carbon, combusted, absorbed and analyzed by IC.
PFHxS	Perfluorinated sulfonate standard	Sodium perfluoro-1-hexanesulfonate standard. Used for MDL, and QC check standards.
RSE	Relative Standard Error	Error from calibration line. Defines the best fit of a calibration curve.
TSS	Total Suspended Solids	Includes any loose particulates or precipitated solids in the water sample.

 Learn more at thermofisher.com/combustionIC

General Laboratory Equipment – Not For Diagnostic Procedures. © 2025 Thermo Fisher Scientific Inc.

All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified. Hamilton is a trademark of Hamilton Co. SPEX CertiPrep is a trademark of SPEX CERTIPREP, Inc. Ricca Chemical is a trademark of RICCA Chemical company. MilliporeSigma, Supelco, and TraceCERT are trademarks of Merck KGaA, Darmstadt, Germany and/or its affiliates. ASTM is a trademark of the American Society for Testing and Materials. Teflon is a trademark of Chemours Company. Nittoseiko, ASC-270LS and AQF-2100H are trademarks of Nittoseiko Analytech Co., Ltd. This information is presented as an example of the capabilities of Thermo Fisher Scientific products. It is not intended to encourage use of these products in any manner that might infringe the intellectual property rights of others. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details. **AN002748-EN 0125S**